



Ref. Certif. No.

**BE-41344**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

### CB TEST CERTIFICATE

Product

Robotic component for medical device

Name and address of the applicant

KUKA Deutschland GmbH  
Zugspitzstr. 140, 86165, Augsburg, Germany

Name and address of the manufacturer

KUKA Deutschland GmbH  
Zugspitzstr. 140, 86165, Augsburg, Germany

Name and address of the factory

See page 2

Additional Information on page 2

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

Sunrise Cabinet Med: 230/110 V AC; 50/60 Hz, 4,7/10 A;  
LBR Med Robot: Supplied by Sunrise Cabinet Med via coded connecting cable

Trademark (if any)

KUKA

Customer's Testing Facility (CTF) Stage used

-

Model / Type Ref.

Sunrise Cabinet Med, LBR Med 7 R800, LBR Med 14 R820

Additional information (if necessary may also be reported on page 2)

-  
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020

National Differences:

CA, US, CH

S26H0001


As shown in the Test Report Ref. No. which forms part of this Certificate

This CB Test Certificate is issued by the National Certification Body

SGS Belgium NV - Division SGS CEBEC  
Riverside Business Park  
Bld Internationaalelaan 55, Building K  
B-1070 Brussels, Belgium



Date: 2022-04-05

  
Signature: Calogero Lana



Ref. Certif. No.

**BE-41344**

**Name and address of the factories:**

1. KUKA Deutschland GmbH  
Zugspitzstr. 140, 86165, Augsburg, Germany
2. KUKA Robotics Hungaria Ipari Kft. Fűzesgyarmat  
Pozsonyi u. 15, 5525, Fűzesgyarmat, Hungary

**Additional information:**

-

SGS Belgium NV - Division SGS CEBEC  
Riverside Business Park  
Bld Internationalelaan 55, Building K  
B-1070 Brussels, Belgium



Date: 2022-04-05

  
Signature: Calogero Lana